



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
REGION 10 LABORATORY
7411 Beach Dr. East
PortOrchard, Washington 98366

MEMORANDUM

DATE: March 3, 2009
TO: Rob Pedersen, Project Officer, OEA

FROM: Stephanie Harris, Technical Director, Microbiology
Region 10 Laboratory

SUBJECT: Final report for the Micro Drysuit Decon
Project Code: LAB 502R
Account Code: 0809B10P202BD4C

cc: Barry Pepich, Laboratory Director
USEPA Region 10

The following is a final report discussing the results for the Drysuit Decon Study using *Pseudomonas aeruginosa*. The analyses were performed following the Standard Operating Procedures (SOP) identified in the associated Quality Assurance Project Plan (QAPP) developed by OEA with the assistance of the USEPA Region 10 Laboratory. This report is relevant for the following samples:

Sample numbers 08290300 – 0317; 08300300 – 0317.

1.0 Sample Analysis and Determination of Results

- 1.1 6" x 6" squares of dry suit diving material were disinfected and allowed to air dry prior to use in this study.
- 1.2 A standard aliquot of *Pseudomonas aeruginosa* containing approximately 10,000 cfu was applied to the swatches of pre-cleaned dive suit material. The level of bacteria in the standard aliquot was determined directly with each separate analytical run by testing in triplicate. Data table refers to column A (D2).
- 1.3 The number of organisms recovered from the dive suit material after a 3 or 1 minute contact time on the dive suit material in the absence of the disinfectant was used as the bacterial load in the applied aliquot, or the "results per 100 ml." This was also done in triplicate and results were averaged to obtain the level demonstrated in the report.
- 1.4 Triplicate samples of dry suit material were used to determine the effectiveness of a 4.7 % solution of betadine at reducing the number of bacteria present on the dive material within a timed exposure period. This was also done in triplicate for each set of data. Data table refers to log removal and percent removal.

2.0 Quality Control Tests Performed:

As established in the Quality Assurance Project Plan for this project, the following quality control tests were conducted as an integral part of these analyses:

- 2.1 Negative control – triplicate sets of pre-cleaned dive suit material was rinsed with buffer and the residue collected and filtered to check for background *Pseudomonas* contamination. A positive result (growth) from a negative control would have invalidated the data associated with that set.

Negative filtration control –Filtration of 100 ml of sterile rinse water performed to ensure that the filtration portion of the analysis demonstrates no bacterial contamination. Filters are placed on media and incubated with the test samples. A positive result (growth) on this control invalidates the data associated with the set.

Positive control – Standard aliquots of *Pseudomonas* culture were applied to triplicate swatches of dive suit material. After the timed interval without exposure to the disinfectant, the dive suit material was rinsed; with the rinsate being directly filtered through 47 mm diameter, 0.45 µm porosity filters. The filters were placed on mPA agar and incubated. The number of organisms counted and factoring in the dilution used, the number obtained was used in calculations determining the percent removal or log removal of organisms during the disinfection step. A negative result (no growth) on this control would have invalidated the data associated with this set.

3.0 General Conclusions and Disclaimers:

3.1 For this study, two time exposure intervals were utilized to determine effectiveness of the disinfectant at reducing the levels of viable organisms. Three minutes was used initially, but once it was realized that essentially 100 % of organisms were rendered nonviable at three minutes, the study was expanded to include one minute exposures. Although the results were somewhat lower for the three minute exposure (more organisms viable), both time intervals resulted in greater than 3 log (99.9 %) reduction in viable organisms and the results are not statistically significant.

3.2 Objectives to study:

- 1) **Determine the efficacy of the potable water rinse procedure for removal of bacteria on the Viking drysuit material.**
 - a. By comparison of column D2 (initial seeding) and column C in all sets of data, it is apparent that a potable water rinse will effectively remove up to 100 % of the organisms, although there is a range of removal from 74 – 100 % (rsd 13.3 %). However, the organisms rinsed off the dive suit were viable and could present a biological hazard on board the vessel or in the waterway.
- 2) **Determine the efficacy of Betadine to kill bacteria present on the Viking drysuit material. Two different application periods (1 minute and 3 minutes) will be used.**
 - a. The data compiled from 3 sets of triplicate studies demonstrated a 99.98 % (rsd = 0.15 %) reduction of viable organisms using 4.7 % betadine solution and a 3 minute exposure time.
 - b. The data compiled from 3 sets of triplicate studies demonstrated a 99.99 % (rsd = 0.006 %) reduction of viable organisms using a 4.7 % betadine solution and a 1 minute exposure time.
- 3) **Determine the efficacy of hydrogen peroxide to kill bacteria present on the Viking drysuit materials. Two different application periods (1 minute and 3 minutes) will be used.**
 - a. Not yet completed.

4) **The primary intent is to remove bacteria from the suit; absent this, the intent is to kill bacteria in-situ.**

- a. See discussion in section 1 and 2 above. Although the organisms are readily removed from the dive suit using a potable water rinse, the organisms are viable at this point and could represent a safety risk or a source of contamination to the water way.

Abbreviations used in data report:

- 3.7.1 ND - Analysis Not Done
- 3.7.2 Cfu – colony forming units (basically number of bacteria, assuming a single colony represents a single bacteria)

**Quality Assurance Project Plan
for
Viking Dive Suit Decontamination Study
EPA Region 10**

**Prepared by
Environmental Services Unit
Office of Environmental Assessment**

July 9, 2008

_____/s/_____
Dr. Stephanie Harris, EPA Region 10/MEL Sr. Microbiologist/Project Lead

_____/7/9/08_____
Date

_____/s/_____
Rob Pedersen, EPA Region 10 Project Officer

_____/7/9/08_____
Date

_____/s/_____
Ginna Greppo-Grove, EPA Region 10 Acting QA Manager

_____/7/9/08_____
Date

Keven McDermott, EPA Region 10 ESU Manager

Date

Table of Contents

<u>Section</u>	<u>Page</u>
1.0 Project Management	4
1.1 Distribution List.....	4
1.2 Project Organization	4
1.3 Problem Definition/Background.....	5
1.4 Project/Task Description and Schedule.....	8
1.5 Data Quality Objectives and Criteria.....	8
1.5.1 Objectives and Project Decisions	8
1.6 Special Training and Certification.....	9
1.7 Documents and Records	9
1.7.1 QAPP Distribution.....	9
1.7.2 Determination Levels.....	9
1.7.3 Measurement Performance Criteria/Acceptance Criteria	9
1.7.4 Laboratory Documentation and Records	9
1.7.5 Quarterly and/or Final Reports	9
2.0 Data Generation and Acquisition.....	9
2.1 Sampling Design (Experimental Design)	9
2.2 Laboratory Procedures.....	10
2.2.1 Health and Safety.....	10
2.3 Analytical Methods - Standard Methods (APHA) 9213E (Membrane Filtration Method for <i>Pseudomonas aeruginosa</i>	11
2.4 Quality Control (QC).....	11
2.4.1 Samples Analyzed by the EPA Region 10 Laboratory.....	11
Replicate Analysis (Analyst and Method Precision):.....	11
Method Accuracy:.....	11
Negative culture controls (non-target organism) will be analyzed on a daily basis to ensure that the media used for the evaluation is appropriately restrictive for growth of non-target organisms.	
Positive culture control organisms (target organism) will be analyzed on a daily basis to ensure that the media is appropriate for the organism's growth.	11
2.5 Instrument/Equipment Testing, Inspection, and Maintenance	11
2.5.1 Field Measurement Instruments/Equipment.....	11
2.5.2 Laboratory Analysis Instruments/Equipment	11
2.6 Instrument/Equipment Calibration and Frequency.....	11
2.6.1 Field Measurement Instruments/Equipment.....	11
2.6.2 Laboratory Analysis Instruments/Equipment	11
2.7 Inspection/Acceptance Requirements for Supplies and Consumables	11
2.7.2 Laboratory Analyses Supplies and Consumables.....	11
2.8 Data Acquisition Requirements (Non-Direct Measurements).....	12
Not applicable to this project.....	12
2.9 Data Management	12
3.0 Assessment and Oversight.....	12
3.1 Assessments/Oversight and Response Actions.....	12
3.2 Reports to Management	12
4.0 Data Review and Usability	12
4.1 Data Review, Verification, and Validation Requirements	12
4.1.1 Data Verification/Peer Review	12
4.1.2 Data Validation	12

4.2 Verification and Validation Methods	12
4.2.1 Data Verification.....	13
4.2.2 Data Validation	13
4.3 Reconciliation with User Requirements	13
APPENDICES	14
APPENDIX A. DATA QUALITY PROCEDURES	15
Figure 1-1. Organization.....	15
Table A-1. Data Quality Objectives Summary	15
Table A-2. Analytical Parameters and Target Limits	16
Table A-5.2 Quality Control Requirements for Analyses	17
Appendix A-3. Data Quality Forms.....	18
A-3.1. Attachment 1 – Sample Alteration Form.....	18
A-3.2. Attachment 2 – Corrective Action Form	19
APPENDIX B. FIELD DOCUMENTATION.....	20
Appendix B-1. Equipment/Instrument Manual	20
Appendix B-2. Standard Operating Procedures.....	20
APPENDIX C. LABORATORY DOCUMENTATION	20
Appendix C-1. QA Manual.....	20
APPENDIX D. DATA EVALUATION.....	21
Appendix D-1. Data Evaluation/Documentation Form.	21
D-1.1. Microbiology Laboratory Data Review/Release Form	21

1.0 Project Management

1.1 Distribution List

Name: Rob Pedersen
Title: EPA Region 10 Unit Diving Officer, Project Officer
Organization: EPA Region 10
Contact Information:
Email: Pedersen.rob@epa.gov

Name: Stephanie Harris
Title: EPA Region 10 Laboratory Microbiologist
Organization: EPA Region 10
Contact Information: 7411 Beach Drive East, Port Orchard, WA 98366.
Telephone: (370) 871-8710, E-mail: Harris.Stephanie@epa.gov

Name: Ginna Grepo-Grove
Title: Acting QA Manager
Organization: EPA Region 10
Contact Information: grepo-grove.gina@epa.gov

Name: Keven McDermott
Title: ESU Manager
Organization: EPA Region 10
Contact Information: mcdermott.keven@epa.gov

Name: Sean Sheldrake
Title: EPA Region 10 Deputy UDO
Organization: EPA Region 10
Contact Information:
Email: Sheldrake.sean@epa.gov

Name: Bruce Duncan
Title: EPA Region 10 Deputy UDO
Organization: EPA Region 10
Contact Information:
Email: Duncan.bruce@epa.gov

1.2 Project Organization

(EPA QA/R-5 A4)

The following individuals are responsible for design and implementation of this project, and/or will be the primary data users and decision makers:

- **Stephanie Harris**, (360) 871-8710, EPA Region 10 Environmental Laboratory investigator, is responsible for assisting with the preparation of the quality assurance project plan (QAPP), analysis of samples, and preparation of the final report.
- **Rob Pedersen**, (206) 553-1646, Project Manager, will serve as the primary point of contact for the project.

1.3 Problem Definition/Background

(EPA QA/R-5 A5)

1.3.1. Background

The EPA Region 10 dive team members work in polluted water conditions – contaminated water diving (CWD) from both biological and chemical hazards. Examples are diving near wastewater treatment plant outfalls, storm water outfalls, seafood processor waste, and general chemical contaminants in the water column and sediment. The divers' suits are made of vulcanized rubber for easier decontamination (Viking Pro EPDM/natural rubber (1000 gr/m²).

Decontamination is intended not only to rid the diver of bottom sediments, but to dilute and rinse off microbial contamination; Reference: EPA Diving Safety Manual, Appendix L. Diver dress is typically an AGA full face mask (mated to the hood of the viking drysuit), dry gloves, a Viking drysuit, a buoyancy compensator, and SCUBA equipment. See

<http://yosemite.epa.gov/r10/oea.nsf/webpage/dive+team>, Equipment page for more details. This diver dress renders the diver completely dry during contaminated water diving, absent suit punctures or fit issues with the full face mask. Though the improperly equipped diver with a wetsuit and bite mouth regulator is exposed during the dive itself, EPA diver exposure during a dive operation would be more likely after doffing contaminated equipment that has been inadequately decontaminated. Investigation as to the adequacy of potable water rinse as the day to day decontamination technique is needed to determine whether this is sufficient, or if more elaborate antimicrobial decontamination is worth the additional time and effort to employ. Antimicrobial solutions take significantly more effort to deploy with SCUBA equipment due to the limited air supply carried by the diver. If the diver must spend a significant amount of time performing decontamination, this is time that cannot be spent doing work diving. Time may also become a hazard if the diver is overheating in the sun.

Many heavily contaminated water diving operations are conducted with surface supply to provide unlimited air for decontamination. However, surface supply operations severely limit the work the diver can undertake without relocating the dive platform a number of times. For long transect dive surveys, use of a surface supply system therefore becomes impractical. For these reasons, it is in the divers' interest to have decontamination that is both very effective and of the shortest duration possible.

Decontamination of divers occurs as they exit the water onto the swim step on the stern of the vessel (exclusion zone). Decon generally consists of a thorough potable water rinse. If exposure occurred with a known high biological hazard, then the divers may be sprayed with a diluted Betadine solution (providone-iodine) as a disinfectant. Betadine is the name of [Purdue Pharma's](#) brand of consumer-available [povidone-iodine](#) (PVPI) [topical antiseptics](#). Betadine, is available as a solution, sold [over-the-counter \(OTC\)](#) for cleaning minor wounds^[1] and used in hospitals to prepare a patient's skin prior to surgery.^[2] Solutions are 10% providone-iodine in water. A contact time of two minutes is recommended for some disinfection uses with the Purdue Pharam's product. The Betadine solution used by the dive

team for soaking AGA masks and for Viking suit decon-spray is 9 oz. Betadine to 1.5 gal. freshwater (a 4.7 percent solution).

A common disinfectant is diluted chlorine bleach. This substance is corrosive to silicone and vulcanized rubber as is not used by the dive team. Hydrogen peroxide (H₂O₂) may be an effective disinfectant for use on the Viking suits. The common household hydrogen peroxide is ten percent by volume. Information on hydrogen peroxide:

10% (v/v) Aqueous Solution (Approximately 3% by weight)

Corrosiveness of hydrogen peroxide

The corrosiveness of process water due to hydrogen peroxide depends on the amount of dissolved oxygen that is produced. Oxygen corrodes iron-containing metals. The amount of iron and the pH are a greater influence on corrosiveness than the concentration of hydrogen peroxide is.

How is hydrogen peroxide transported and stored?

Hydrogen peroxide must be transported in polyethylene, [stainless steel](#) or [aluminium](#) containers. When hydrogen peroxide comes in contact with flammable substances, such as wood, paper, oil or cotton (cellulose), spontaneous ignition may occur. When hydrogen peroxide is mixed with organic matter, such as alcohols, acetone and other ketones, aldehydes and glycerol, heavy explosions may occur. When hydrogen peroxide comes in contact with substances, such as [iron](#), [copper](#), chromium, [lead](#), [silver](#), [manganese](#), [sodium](#), [potassium](#), [magnesium](#), [nickel](#), [gold](#), [platinum](#), metalloids, metal oxides or metal salts, this may result in powerful explosions. This is why hydrogen peroxide is usually transported in diluted form.

Is hydrogen peroxide efficient?

The efficiency of hydrogen peroxide depends on several factors, such as pH, catalysers, temperature, peroxide concentration and reaction time.

For reactivity to Viking suit material, the manufacture's chemical permeation test results were consulted.

From Viking Dry Suits CD-ROM:

Table#2: Chemical Permeation Test Results on

Chemical	Concentration %	Solubility in Water	Specific Gravity	Breakthrough Time Minutes		
				PRO	HD	Latex
Acetone	10	100	0.79	50	60	90
Acetonitrile	10	100	0.78	>180	>180	40
Ammonia Solution	10	100	*	>180	>180	>180
Carbon Disulphide	100	0.2	1.26	1	1	8
Dichloromethane	100	1.3	1.34	5	5	17
Diethylamine	10	82	0.71	>180	>180	>180
Dimethylformamide	10	100	0.95	>180	>180	>180
Ethyl Acetate	8.7	8.7	0.9	20	52	65
Hexane	0.014	0.014	0.66	>180	>180	>180
Methanol	10	100	0.79	>180	>180	>180
Sodium Hydroxide	10	50	2.13	>180	>180	>180
Sulphuric Acid	10	100	1.83	>180	>180	>180
Tetrachloroethylene	0.015	0.015	1.62	>180	>180	40
Tetrahydrofuran	10	100	0.89	60	>180	80
Toluene	0.05	Not soluble	0.87	>180	>180	85
Oil No. 1 acc. to ISO 1817**	100	Not soluble	<1	>180	>180	>180

Sodium hydroxide is an oxidizer similar to hydrogen peroxide. The table above shows a breakthrough time of greater than three hours in a ten percent solution of NaO₂. There was no effect on the suit seams as shown in the table below.

Table #3: Diffusion through Seams for Viking Suits

Chemical	Concentration	Solubility in Water	Specific Gravity	Diff G/HR	Effect
Acetone	10	100	0.79	4.8	No effect
Acetonitrile	10	100	0.78	0.02	No effect
Ammonia Solution	10	100	*	0.01	No effect
Dichloromethane	100	1.3	1.34	48	Reversible swell
Diethylamine	10	82	0.71	0.36	No effect
Dimethylformamide	10	100	0.95	0.02	No effect
Ethyl Acetate	8.7	8.7	0.9	0.01	No effect
Ethyl Acetate**	100	8.7	0.9	4.1	Low reversible swell
Hexane	0.014	0.014	0.66	0.03	No effect
Hexane**	100	0.014	0.66	29	Moderate reversible swell
Methanol	10	100	0.79	0.25	No effect
Sodium Hydroxide	10	50	2.13	0.01	No effect
Sulphuric Acid	10	100	1.83	0.02	No effect
Tetrachloroethylene	0.015	0.015	1.62	0.03	No effect
Tetrahydrofuran	10	100	0.89	0.05	Reversible swell
Tetrahydrofuran**	100	100	0.89	27.5	Seam starts to delaminate
Toluene	.05	.05	0.87	0.04	No effect

The required contact time for Betadine to be an effective disinfectant under the conditions the dive team experiences on the back of the EPA vessel is not well known. For other topical disinfectant applications with Betadine, contact times of up to three minutes are mentioned. It may also be necessary for the Betadine solution to dry after application.

Given the reaction of hydrogen peroxide to biological substances, a relatively short contact time may be sufficient. Also, drying in-place on the diver's suit may not be necessary for hydrogen peroxide. The diver team has not used hydrogen peroxide for disinfection purposes.

This decon study will compare the effectiveness of freshwater rinsing of Viking suit material to rinsing suit material after a one minute and a three minute exposure to Betadine, and after a one minute and a three minute exposure to ten percent hydrogen peroxide.

1.3.2 Objectives and Goals

The objectives of the decontamination procedure study are to:

- 1) Determine the efficacy of the potable water rinse procedure for removal of bacteria on the Viking drysuit material.
- 2) Determine the efficacy of Betadine to kill bacteria present on the Viking drysuit material. Two different application periods (1 minute and 3 minutes) will be used.
- 3) Determine the efficacy of hydrogen peroxide to kill bacteria present on the Viking drysuit materials. Two different application periods (1 minute and 3 minutes) will be used.
- 4) The primary intent is to remove bacteria from the suit; absent this, the intent is to kill bacteria in-situ.

Information from this study will be used to modify diver decontamination procedures. Results will also be presented to the EPA National Diving Safety Board and peer groups.

1.4 Project/Task Description and Schedule

This QAPP provides the supportive information used in developing the study plan for laboratory methods testing of diver decontamination procedures.

Project Schedule

- 1) Develop methods plan – June 2008.
- 2) Obtain culture media and surrogate bacteria species *Pseudomonas aeruginosa*. – June 2008.
- 3) Perform bacterial test growths. – June 2008.
- 4) Perform decontamination study. – July 2008.
- 5) Analysis. – July - August 2008.
- 6) Study report. – October 2008.

1.5 Data Quality Objectives and Criteria

Data from this phase of the project will be used as background to develop further data quality objectives for this project.

1.5.1 Objectives and Project Decisions

The primary data quality objective for the diver suit microbial source decontamination study is to characterize the effectiveness of three decontamination procedures. If a potable water rinse is effective at removing relatively high concentrations of bacteria on Viking suit material, then antimicrobial use

may be unnecessary for the majority of Region 10 CWD exposures (thought to be at relatively low concentrations of bacteria).

1.6 Special Training and Certification

The Region 10 Environmental Laboratory staff has completed the required annual 8-hour health and safety training. The Region 10 Laboratory is accredited by the National Environmental Laboratory Accreditation Council. The Region 10 Environmental Laboratory analysts have the appropriate training to conduct bacteriological analyses.

1.7 Documents and Records

1.7.1 QAPP Distribution

It will be the responsibility of the project manager to ensure that appropriate project personnel have the most current, approved version of the QAPP, including updates. The final version of the QAPP and any updates will be distributed in portable document file format.

1.7.2 Determination Levels

Concentrations of inocula to be placed on the Viking drysuit material to test the efficacy of the disinfectant will be determined by previous serial dilution testing. Fresh cultures of *Pseudomonas aeruginosa* (ATCC 27853) will be used each test day. A minimum of 10^4 organisms will be placed on the suit patch in order to ensure that a 99.0 % (or 2 log) removal can be determined.

Method detection limit is 1 CFU (colony forming unit) per 100 ml.

1.7.3 Measurement Performance Criteria/Acceptance Criteria

The measurement performance criteria/acceptance criteria for this project are discussed in Section 2.4, Quality Control. In general, if a sample, or associated controls, fall outside of the acceptance criteria, they are invalidated and either re-sampled or re-analyzed, as appropriate.

1.7.4 Laboratory Documentation and Records

Laboratory documentation will include but is not limited to raw data, sample preparation and analysis logs, and results of calibration and quality control (QC) checks.

1.7.5 Quarterly and/or Final Reports

The EPA Region 10 Laboratory will archive the electronic and hard copies of analytical data for a minimum of 10 years as per the QA manual.

2.0 Data Generation and Acquisition

The elements in Sections 2.1-2.10 ensure that appropriate methods for sampling, measurement and analysis, data collection, data handling, and quality control (QC) activities are employed and documented.

2.1 Sampling Design (Experimental Design)

This outline of the laboratory procedure will occur twice, once with Betadine and once with hydrogen peroxide:

2.1.1 Use 6"X6" patches of Viking suit material.

2.1.2 Clean the patches (soap/water, disinfect, air dry).

2.1.3 4 patches placed on a clean surface.

2.1.3.1 Patch 1 inoculated by swabbing with *Pseudomonas aeruginosa* (control patch -this will also catch any potential background that remains on the patches).

2.1.3.2 Patches 2-4 inoculated by swabbing with *Pseudomonas aeruginosa* (will have a known concentration of *Pseudomonas* - this is part of the purpose of patch 1).

2.1.3.3 Allow patches to partially dry (5 minutes), but not completely.

2.1.4 Hang patches over clean pans or beakers.

2.1.5 Decon:

2.1.5.1 Patches 1-2, rinse with sterile water (patch 1 will be the control and patch 2 a “test” for sterile water rinsing as appropriate for removing the bacteria).

2.1.5.2 Patches 3-4, spray with Betadine solution (dilution ratio is 9 oz. Betadine to 1.5 gal. freshwater) for decon test one, spray with ten percent hydrogen peroxide for test two.

2.1.5.3 Patch 3, rinse with sterile water after 1 min. of Betadine/or H₂O₂ exposure.

2.1.5.4 Patch 4, rinse with sterile water after 3 min. of Betadine/ or H₂O₂ exposure.

2.1.6 Filter a volume of the rinsates (usually 100 - 500 ml) through a 0.45 um porosity 47 mm diameter membrane filter.

2.1.7 After rinsing the filtration funnel, apply the membrane on a poured plate of mPAC agar so that there is no air space between the membrane and the agar surface.

2.1.8 Invert plates and incubate at 41.5 +/- 0.5 degree C for 72 hours.

Media - m PA - C (modified m-PA agar) available commercially. Media will be prepared to manufacturer's requirements.

2.1.9 Count colonies and report as number of *P. aeruginosa*/100 ml.

2.1.10 Determine percent “kill” in bacteria count (on a log basis) in rinsates 2-4 relative to rinsate 1, and between rinsate 2 with rinsates 3-4.

Note: there will be up-front work to determine the amount of inoculum to place on the patches. In order to have log removals, it is necessary to start with large numbers of organisms and then perform dilutions with the inoculum to obtain countable numbers.

Methods reference: Standard Methods for Examination of Water and Wastewater, 21st edition. 9213E, Membrane Filter Technique for *Pseudomonas aeruginosa*.

2.1.11 Repeat the decon test at least twice more, on different days.

2.1.12 Determine if there is a statistically significant difference between rinsates 1-2 with rinsate 3 and rinsate 4 (and also between 3 and 4). If there are an adequate number of samples, the analysis method might be a series of paired T-tests (for homoscedastic data) or multivariate analysis for non-normal data.

2.2 Laboratory Procedures

Analytical methods, expected range of results, and required detection limits are summarized in Table 4.

Parameter	Description	Method	Lab	Sample Container	Preservation	Holding Time	Precision/ Quantitation Limits
<i>Pseudomonas</i>	Membrane filtration	APHA 9213E	EPA	NA	NA	8 hours	20% RSD*/ 1 cfu/100 mL

* RSD-Relative standard deviation, standard deviation divided by the mean

Table 4 analytical Methods Summary

2.2.1 Health and Safety

When working with potentially hazardous materials, investigators are to follow USEPA, Occupational Safety and Health Administration and site-specific health and safety procedures.

2.3 Analytical Methods - Standard Methods (APHA) 9213E (Membrane Filtration Method for *Pseudomonas aeruginosa*).

2.4 Quality Control (QC)

The following QC activities will be performed by the laboratories performing analytical services in support of this project.

2.4.1 Samples Analyzed by the EPA Region 10 Laboratory

Replicate Analysis (Analyst and Method Precision):

Replicate analyses will occur with 10 % of laboratory generated samples. Analyst precision will be determined by duplicate counts of the same plate. Analysts must be within 10 % for two analyst counts and 5 % for single analyst counts.

Method Accuracy:

Negative culture controls (non-target organism) will be analyzed on a daily basis to ensure that the media used for the evaluation is appropriately restrictive for growth of non-target organisms. Positive culture control organisms (target organism) will be analyzed on a daily basis to ensure that the media is appropriate for the organism's growth.

2.5 Instrument/Equipment Testing, Inspection, and Maintenance

2.5.1 Field Measurement Instruments/Equipment

This phase of the project requires no field instruments.

2.5.2 Laboratory Analysis Instruments/Equipment

Laboratory instruments such as incubators, pH meters and other equipment required by the applicable analytical methods will be maintained according to the manufacturers' instructions and the Laboratory SOPs. Records for equipment service shall be maintained by the Laboratory.

2.6 Instrument/Equipment Calibration and Frequency

2.6.1 Field Measurement Instruments/Equipment

No field instruments will be used during this phase of the project.

2.6.2 Laboratory Analysis Instruments/Equipment

Laboratory equipment (e.g. pH meter, incubator, etc.) will be calibrated using the method and frequency specified in the Laboratory's SOPs. Records on calibration of laboratory equipment shall be maintained by the Laboratory.

2.7 Inspection/Acceptance Requirements for Supplies and Consumables

2.7.2 Laboratory Analyses Supplies and Consumables

The quality of chemicals, media and other supplies and consumables used in the Laboratory is dictated by the sensitivity and specificity of the analytical techniques being used. In the Region 10 Laboratory, chemicals, media, and other supplies and consumables are marked with the date received and the receiver's initials. In the event an expiration date has not been assigned by the manufacturer, the

expiration date will be assigned by the receiver according to the Region 10 Laboratory's work instruction, "General Guidelines for Assigning Standard Expiration Dates." The quality of all laboratory supplies is documented by the supplier and the Region 10 Laboratory requests and keeps the vendor certificates on record per NELAC requirements.

2.8 Data Acquisition Requirements (Non-Direct Measurements)

Not applicable to this project.

2.9 Data Management

The Laboratory will maintain a logbook that includes the time of analysis and analyst initials. Quality control results will be recorded on bench sheets. All data generated by Region 10 will be subject to a peer review then signed-off by the Microbiology Team Technical Director. Data entry staff will process and distribute all information mentioned above in accordance with the Laboratory's SOP. Logbooks, bench sheets and final reports will be stored on-site. All data generated during this project will be processed, stored, and distributed according to Laboratory SOPs.

3.0 Assessment and Oversight

3.1 Assessments/Oversight and Response Actions

Laboratories routinely perform performance checks using method-specific positive and negative controls, blind samples, etc.

Corrective actions will be implemented in response to any QA results or detection of unacceptable data.

These corrective actions will be developed in consultation with the Office of Research and Development, keeping the data user informed of any impacts on the data. If required, corrective actions will be documented in Appendix A-3.2, Corrective Action Form.

3.2 Reports to Management

A final report will be generated at the completion of the project. This report will include a discussion of the findings, interpretation of data and an executive summary. The report will be provided to all individuals listed in section 1.1 (Distribution List).

4.0 Data Review and Usability

4.1 Data Review, Verification, and Validation Requirements

4.1.1 Data Verification/Peer Review

Region 10 data verification and peer review will be accomplished following the Laboratory's SOP Mi_D001A (Data Review). Data will be qualified as necessary to convey to the user any important information that needs to be considered in its use.

4.1.2 Data Validation

Data validation is an evaluation of the technical usability of the verified data with respect to the planned objectives of the project. This is accomplished by applying a defined set of performance criteria to the body of data in the evaluation process. Data validation for this project will be performed by the project manager.

4.2 Verification and Validation Methods

4.2.1 Data Verification

Verification of the Region 10 analytical results is the responsibility of the Microbiology Technical Director, as required by the Laboratory's QA Manual. If any deviations are identified, the potential impact of those deviations on the reliability of the data will be assessed, and the information will be provided to the project manager through the QA Memo and appropriate flagging of the data.

4.2.2 Data Validation

Data validation will evaluate all individual samples collected and analyzed to determine if the results are within acceptable limits. Quantitative or qualitative limits of acceptability are defined for precision, accuracy, representativeness, comparability, and completeness.

- 1) Precision is defined as the agreement between a set of replicate measurements without assumption and knowledge of the true value. Agreement is expressed as either the relative percent difference (RPD) for duplicate measurements or the range and standard deviation for larger numbers of replicates. Data on precision are obtained by analyzing duplicate and replicate samples.
- 2) Accuracy is a measure of the closeness of a sample analysis result to the "true" value. Accuracy will be determined primarily by an evaluation of the agreement between repeat analyses within the laboratory.

Representativeness is defined as the degree to which data accurately and precisely represents characteristics of a population, parameter variations at a sampling point, or an environmental condition. For this project, representativeness will be ensured by selection of diver suit material and surrogate decontamination techniques in accordance with the sampling design requirements in this QAPP.

- 3) Data are comparable if collection techniques, measurement procedures, methods, and reporting units are equivalent for the samples within a sample set. Comparable data for this project will be obtained by specifying standard units and using standard procedures for sample processing and analysis.
- 4) Data are complete when a prescribed percentage of the total intended measurements and samples are obtained. Analytical completeness is defined as the percentage of valid analytical results requested. For this project, acceptable completeness is > 90%.

4.3 Reconciliation with User Requirements

All data and related information obtained during the course of this project will be included in a data report. Presentations of data and data analysis may be made to relevant user groups upon request.

APPENDICES

Appendix A.

Appendix C. Laboratory Documentation

C-1. QA Manual

C-2. Standard Operating Procedures

C-3. Data Report Forms

Appendix D. Data Evaluation

APPENDIX A. DATA QUALITY PROCEDURES

Figure 1-1. Organization

Project Organization: Refer to section 1.2 for project participant roles and responsibilities.

Table A-1. Data Quality Objectives Summary

Analytical Group	Number of Samples	# of QA Samples: Reference Samples	Matrix	Method	Method Detection Limits	Accuracy	Precision (RPD)	Completeness	Volume, Container	Holdin g Time (days)
Laboratory Measurements										
<i>Pseudomonas aeruginosa</i>	4 per event	1 each batch	Water	Standard Methods (9213E)	1 CFU	+/- 10%	+/- 10%	> 90%	clean dry suit	6 hours

1 – Standard Accuracy and Precision for analysis by PCR is unknown at this time. Identification is not quantitative.
 CFU: Colony Forming Unit

Table A-2. Analytical Parameters and Target Limits

Matrix/Media:

Analytical Parameter	Project Action Limit/Level (applicable units)	Laboratory Limits ¹ (applicable units)	
		Quantitation Limits	Detection Limits (if appropriate)
<i>Pseudomonas aeruginosa</i>	1 cfu/100 mL	1 cfu/100 mL	1 cfu/100 mL

¹ Laboratory quantitation limits and detection limits are those that an individual laboratory or organization is able to achieve for a given analysis on a routine basis.

- Quantitation limits are the minimum concentrations that can be identified and quantified above the detection limit within some known limits of precision and accuracy/bias. It is recommended that the quantitation limit is supported by the analysis of a standard of equivalent concentration (typically, the lowest calibration standard).
- Detection limits are the minimum concentration that can be detected above background or baseline/signal noise of an instrument.

cfu: colony forming unit

Table A-5.2 Quality Control Requirements for AnalysesAnalytical Method/SOP: PCR qualitative²

QC Sample:	Data Quality Indicator (DQI)	Frequency/Number	Method/SOP QC Acceptance Limits	Acceptance Criteria/Measurement Performance Criteria ¹	Corrective Action
LABORATORY ANALYSIS:					
Filtration control		1 per day	Negative, no DNA	negative, no growth	Data reviewed, decision made based on cause
Positive control		1 per day	Positive reaction	Appropriate media reaction	Data reviewed, decision made based on cause
Negative control		1 per day	Negative reaction	Appropriate media reaction	Data reviewed, decision made based on cause
Replicate analysis or duplicate count		10 % of samples	Within 20 % RSD. 10 % between analysts; 5 % with one analyst	Appropriate DNA amplification	Data reviewed, decision made based on cause
FIELD ANALYSIS:					
Not applicable					

Appendix A-3. Data Quality Forms.

A-3.1. Attachment 1 – Sample Alteration Form

Project Name and Number: _____

Material to be Sampled: _____

Measurement Parameter:

Standard Procedure for Field Collection & Laboratory Analysis (cite reference):

Reason for Change in Field Procedure or Analysis Variation:

Variation from Field or Analytical Procedure:

Special Equipment, Materials or Personnel Required:

Initiators Name: _____ Date: _____

Project Officer: _____ Date: _____

QA Officer: _____ Date: _____

A-3.2. Attachment 2 – Corrective Action Form

Project Name and Number: _____

Sample Dates Involved: _____

Measurement Parameter:

Acceptable Data Range:

Problem Areas Requiring Corrective Action:

Measures Required to Correct Problem:

Means of Detecting Problems and Verifying Correction:

Initiators Name: _____ Date: _____

Project Officer: _____ Date: _____

QA Officer:

APPENDIX B. FIELD DOCUMENTATION

Appendix B-1. Equipment/Instrument Manual

Not applicable to this project.

Appendix B-2. Standard Operating Procedures

Not applicable to this project.

APPENDIX C. LABORATORY DOCUMENTATION

Appendix C-1. QA Manual

Laboratory analysis and procedures will comply with the guidelines described in the document entitled, *Quality Assurance Manual for the U.S. EPA Region 10 Manchester Environmental Laboratory (October 2005)*. The QA Manual is available at the following website on EPA's Intranet (G:\Sops\NELAC 2005 QAM\NELACTable.html). If you are unable to access this document and would like to obtain a copy, please contact Stephanie Harris.

APPENDIX D. DATA EVALUATION

Appendix D-1. Data Evaluation/Documentation Form.

D-1.1. Microbiology Laboratory Data Review/Release Form

Project: _____ Project Code: _____

Sample Numbers _____

Peer Reviewed by: _____

Date: _____

Raw Data/Quality Control Check

- ___ Verify positive and negative culture controls associated with media are satisfactory.
- ___ Verify media sterility was checked.
- ___ Check for sample carryover/contamination if membrane filtration method used. Note any deficiencies.
- ___ Check duplicate analyst counts are within 20 %, when applicable.
- ___ Verify that media was prepared within method specifications.
- ___ Verify that samples were received and analyzed within the holding time.

Bench Sheet Check

- ___ Is the data package properly labeled?
 - ___ Analyst name
 - ___ Sample numbers and project name
 - ___ Analytical method used
 - ___ Date and time of collection/analysis
- ___ Verify that there is a bench sheet for each sample listed on the Analysis Required forms.
- ___ Verify that there is a Data Review Memo written for the project -forwarded to ESAT Data Entry Technician
- ___ Verify that there is a Data Release Memo for this project - forwarded to ESAT Data Entry Technician

Results

- ___ Verify that the reported results:
 - ___ have appropriate qualifiers assigned
 - ___ reflect the correct units
 - ___ reflect dilution factors used in the analysis
 - ___ were transferred correctly from the bench sheets
 - ___ were calculated correctly



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
REGION 10
1200 Sixth Avenue
Seattle, Washington 98101

July 10, 2008

Reply To
Attn Of: **OEA-095**

MEMORANDUM

SUBJECT: Review of Quality Assurance Project Plan for the Viking Dive Suit Decontamination Study, EPA Region 10, July, 2008

FROM: Donald Matheny, Chemist
Office of Environmental Assessment (OEA-095)

TO: Rob Pedersen, Project Officer
Office of Environmental Assessment (OEA-095)

I've completed a review of the above Plan and overall approval is provided. If you have any questions, please call me at (206)553-2599.



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
REGION 10 LABORATORY
7411 Beach Dr. East
Port Orchard, Washington 98366

MEMORANDUM

SUBJECT: Data Release for Microbiology analysis results from the USEPA Region 10 Laboratory.

PROJECT NAME: Micro Drysuit Decon Study

PROJECT CODE: LAB 502R

FROM: Gerald Dodo, Chemistry Supervisor
Office of Environmental Assessment, USEPA Region 10 Laboratory

TO: Rob Pedersen, Project Manager
Office of Environmental Assessment, USEPA Region 10

CC: Barry Pepich, Laboratory Director
Office of Environmental Assessment, USEPA Region 10 Laboratory

I have authorized release of this data package. Attached you will find the microbiology results for the Drysuit Decon Study samples collected 7/15 – 7/24/2008. This is the last of the data associated with this project. For further information regarding the attached data, contact Stephanie Harris at 360-871-8710.



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
REGION 10 LABORATORY
7411 Beach Dr. East
Port Orchard, Washington 98366

DATE: March 3, 2009

TO: Rob Pedersen, Project Officer
Office of Environmental Assessment, USEPA Region 10

FROM: Stephanie Bailey, Microbiologist
Office of Environmental Assessment, USEPA Region 10 Laboratory

SUBJECT: Quality Assurance Review of the Drysuit Decon Study
Project Code: LAB-502R
Account Code: 0809B10P202BD4C

CC: Barry Pepich, Laboratory Director
USEPA Region 10

The following is a quality assurance review for the results from the *Pseudomonas aeruginosa* analyses from the Drysuit Decon Study Project. The analyses were performed by EPA microbiologists at the US EPA Region 10 Laboratory in Port Orchard, WA, following US EPA and Laboratory guidelines.

This review was conducted for the following samples: **08290300 – 0317 and 08300300 – 0317**

1. Data Qualifications

Comments below refer to the quality control specifications outlined in the Laboratory's current Quality Assurance Manual, Standard Operating Procedures (SOPs) and the Quality Assurance Project Plan (QAPP). No excursions were required from the method.

All measures of quality control met Laboratory/QAPP criteria.

The Region 10 Laboratory Quality System has been accredited to the standards of the National Environmental Laboratory Accreditation Conference (NELAC).

2. Sample Transport and Receipt

Sample transport and receipt did not apply.

3. Sample Holding Times

Sample holding times did not apply.

4. Laboratory Quality Control

No qualification was necessary based on laboratory quality control criteria.

All laboratory equipment and supplies used in this analytical procedure met the criteria as set forth in the Standard Operating Procedures (SOP) identified in the associated Quality Assurance Project Plan (QAPP).

All positive and negative control measures demonstrated correct responses for these sets of analyses. No qualifications were required based on quality control measures.

5. Reporting Limits

All sample results that fall below the detection limit are assigned the value of the detection limit with the "<" qualifier attached. No results fell below the detection limit.

6. Changes from Preliminary Data

No changes were made between the preliminary and final data.

7. Data Qualifiers

No qualifications were necessary.

8. Definitions

ND - Analysis not done

cfu – colony forming units (number of bacteria, assuming a single colony represents a single bacteria)



**UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
 REGION 10 LABORATORY
 7411 Beach Dr. East
 Port Orchard, Washington 98366**

**IMPORTANT INFORMATION
 REGARDING ATTACHED FILE**

This file contains data that is readable into Lotus, Excel, WordPerfect, or most databases.

You will need access to PKUNZIP or WINZIP to decompress the file. Once “unzipped” there will be one large file (more appropriate for importing into a database) with the project code as the file name. The fields will be in the following order:

Project ID	Analyte	Matrix
Sample ID	Result	Sample Type Description
Sample Type	Units Code	Sample Description
Parameter Code	Qualifier	Version (Date this file was created)
Analyte Code	Date Collection End	

There will also be multiple smaller files with names such as “METQ1-1.txt,” “GENSA-1.txt,” “BNASA-1.txt,” etc. These files are meant to be imported into Lotus or Excel. To open select File/Open and select file type TEXT or .TXT.

The naming convention is as follows: SSSSTT-#.TXT

Where:

- SSSS: Metals (MET), General (GEN), GCMS(BNA, VOA, BNAT, VOAT), GC(GC)
- TT: Sample Data (SA, Blanks (Q1), Matrix spikes/controls (Q2), Duplicates (Q3)
- #: If the table size exceeds 256 columns then the files will be split into multiple smaller files with sequential numbering. Lotus and Excel can only handle 256 columns.

Sample information appears in the following order:

Sample ID
Sample Description
Sample Type
Matrix
Units

(It will be indicated if a cell contains data of units other than the default.)

Analyte information appears in the following order:

Parameter ID
Method Code
Analyte Code
Analyte Name

For General (also called Classical) and FASP data, sample information appears down the side. All other data has the sample information appearing across the top.

Any questions/suggestions should be e-mailed to Tony Morris at morrris.tony@epa.gov.

08290309	08290310	08290300	08290311	08290301	08290312	08290302	08290313	08290303	08290314	08290304	08290305	08290316	08290306	08290317	08290307	08290318	08290308	08290319	08290309	08290310	08290317	08290307	08300300	08290308	08300301	08300312				
LAB-502R	08300314		08300315		08300316		08300317		08300318		08300319		08300320		08300321		08300322		08300323		08300324		08300325		08300326					
DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES	SPIKED #1 RUN	DI SINFEKTANT 3 MINUTES				
MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE	DI SINFEKTANT 1 MINUTE	MI NUTE				
Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample				
Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample				
Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample	Reg sample				
Sample Data	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d				
Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d	Li qui d				
per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml				
per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml	per 100ml				
PAERU	SM9213E *90248	Pseudomonas aerugi nosa	8000	0	15733	8000	4	15733	24	15733	1	15733	1	15733	9	15733	0	11967	2	11967	2	11967	15	15733	37	15733	51	15733	10200	0
10200	0	10200	1	0	15733	8000	4	15733	24	15733	1	15733	1	15733	9	15733	0	11967	2	11967	2	11967	15	15733	37	15733	51	15733	10200	0

Project ID	Sample ID	Sample Type	Code	Parameter Code	Analyte Code	Analyte Name	Result	Unit	Qualifier	Date Collected	Micro Drysuit Decon Study LAB-502R.txt	Date analyzed	Method Code	Record Version
LAB-502R	08290300	0	PAERU	*90248	Pseudomonas aerugi nosa	8000	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290301	0	PAERU	*90248	Pseudomonas aerugi nosa	8000	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290302	0	PAERU	*90248	Pseudomonas aerugi nosa	8000	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290303	0	PAERU	*90248	Pseudomonas aerugi nosa	1	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290304	0	PAERU	*90248	Pseudomonas aerugi nosa	1	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290305	0	PAERU	*90248	Pseudomonas aerugi nosa	9	per	100ml	Li qui d	7/15/2008	Reg sampl e			
LAB-502R	08290306	0	PAERU	*90248	Pseudomonas aerugi nosa	11967	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290307	0	PAERU	*90248	Pseudomonas aerugi nosa	11967	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290308	0	PAERU	*90248	Pseudomonas aerugi nosa	11967	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290309	0	PAERU	*90248	Pseudomonas aerugi nosa	15	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290310	0	PAERU	*90248	Pseudomonas aerugi nosa	37	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290311	0	PAERU	*90248	Pseudomonas aerugi nosa	51	per	100ml	Li qui d	7/16/2008	Reg sampl e			
LAB-502R	08290312	0	PAERU	*90248	Pseudomonas aerugi nosa	10200	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08290313	0	PAERU	*90248	Pseudomonas aerugi nosa	10200	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08290314	0	PAERU	*90248	Pseudomonas aerugi nosa	10200	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08290315	0	PAERU	*90248	Pseudomonas aerugi nosa	0	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08290316	0	PAERU	*90248	Pseudomonas aerugi nosa	4	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08290317	0	PAERU	*90248	Pseudomonas aerugi nosa	24	per	100ml	Li qui d	7/18/2008	Reg sampl e			
LAB-502R	08300300	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300301	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300302	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300303	0	PAERU	*90248	Pseudomonas aerugi nosa	2	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300304	0	PAERU	*90248	Pseudomonas aerugi nosa	2	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300305	0	PAERU	*90248	Pseudomonas aerugi nosa	1	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300306	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300307	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300308	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300309	0	PAERU	*90248	Pseudomonas aerugi nosa	0	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300310	0	PAERU	*90248	Pseudomonas aerugi nosa	0	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300311	0	PAERU	*90248	Pseudomonas aerugi nosa	1	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300312	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300313	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300314	0	PAERU	*90248	Pseudomonas aerugi nosa	15733	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300315	0	PAERU	*90248	Pseudomonas aerugi nosa	1	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300316	0	PAERU	*90248	Pseudomonas aerugi nosa	5	per	100ml	Li qui d	7/24/2008	Reg sampl e			
LAB-502R	08300317	0	PAERU	*90248	Pseudomonas aerugi nosa	0	per	100ml	Li qui d	7/24/2008	Reg sampl e			

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290300	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #1 RUN			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	8000	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290301	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #1 RUN			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	8000	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290302	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #1 RUN			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	8000	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290303	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	1	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290304	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	1	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/15/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290305	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/15/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	9	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290306	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #2 RUN			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	11967	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290307	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #2 RUN			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	11967	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290308	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #2 RUN			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	11967	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290309	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290310	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	37	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/16/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290311	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/16/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	51	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290312	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #3 RUN			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	10200	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290313	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #3 RUN			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	10200	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290314	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED #3 RUN			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	10200	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290315	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	0	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290316	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	4	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/18/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08290317	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 3 MINUTES			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/18/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	24	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300300	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #1			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300301	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #1			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300302	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #1			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	15733	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300303	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	2	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300304	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	2	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300305	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	1	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300306	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #2			

	Result	Units	Qlfr
--	---------------	--------------	-------------

BIO

Parameter : Pseudomonas aeruginosa	Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Prep Date :

Analytes(s): *90248	Pseudomonas aeruginosa	15733	per 100ml
----------------------------	-------------------------------	--------------	------------------

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300307	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #2			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300308	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #2			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300309	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	0	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300310	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	0	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300311	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
--	---------------	--------------	-------------

BIO

Parameter : Pseudomonas aeruginosa	Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Prep Date :

Analytes(s): *90248	Pseudomonas aeruginosa	1	per 100ml
----------------------------	-------------------------------	----------	------------------

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300312	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #3			

	Result	Units	Qlfr
--	---------------	--------------	-------------

BIO

Parameter : Pseudomonas aeruginosa	Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Prep Date :

Analytes(s): *90248	Pseudomonas aeruginosa	15733	per 100ml
----------------------------	-------------------------------	--------------	------------------

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300313	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #3			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300314	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	SPIKED RUN #3			

		Result	Units	Qlfr
BIO				
Parameter	: Pseudomonas aeruginosa			Container ID : N1
Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method	: SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s):	*90248 Pseudomonas aeruginosa	15733	per 100ml	

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300315	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
--	---------------	--------------	-------------

BIO

Parameter : Pseudomonas aeruginosa	Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)	Prep Date :

Analytes(s): *90248	Pseudomonas aeruginosa	1	per 100ml
----------------------------	-------------------------------	----------	------------------

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300316	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	5	per 100ml

Manchester Environmental Laboratory
Report by Parameter for Project LAB-502R

Project Code:	LAB-502R	Collected:	7/24/08	0:00:00
Project Name:	MICRO DRY SUIT DECON STUDY	Matrix:	Liquid	
Project Officer:	ROB PEDERSON	Sample Number:	08300317	
Account Code:	0910B10P202BD4C	Type:	Reg sample	
Station Description:	DISINFECTANT 1 MINUTE			

	Result	Units	Qlfr
BIO			
Parameter : Pseudomonas aeruginosa			Container ID : N1
Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Analysis Date : 7/24/2008 00:00:00
Prep Method : SM9213E Pseudomonas aeruginosa by membrane filtration (MF)			Prep Date :
Analytes(s): *90248	Pseudomonas aeruginosa	0	per 100ml