

MICHIGAN AIR POLLUTION CONTROL COMMISSION

GENERAL RULES

Effective Date: January 18, 1980

PART 11. CONTINUOUS EMISSION MONITORING

R 336.2153. Zero and drift for continuous emission monitoring systems. (1/18/80)

Rule 1153. (1) The owner or operator of any continuous emission monitoring system required by this part shall do all of the following:

(a) Subject the instruments to the manufacturer's recommended zero and span check at least once daily, unless the manufacturer has recommended adjustments at shorter intervals, in which case such recommendations shall be followed.

(b) Adjust the zero and span whenever the 24-hour zero drift or 24-hour calibration drift limits of the applicable performance specifications in appendix B of 40 C.F.R. part 60 (July 1, 1978), are exceeded.

(c) Adjust continuous monitoring systems purchased prior to September 11, 1974, whenever the 24-hour zero drift or the 24-hour calibration drift exceeds 10% of the applicable emission standard.

(2) Calibration gases used pursuant to subrule (1) shall meet the requirements of rule 1151.