

LabSampleID	AnalysisMethod	Analyte	CommentID	CommentText
0608036-01	VOCs	1,1,2,2-Tetrachloroethane	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	1,1,2,2-Tetrachloroethane	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
0608036-01	VOCs	1,1-Dichloroethene	C1	The reported concentration for this analyte is below the quantitation limit.
0608036-01	VOCs	1,3-Dichlorobenzene	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	1,3-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery
0608036-01	VOCs	1,4-Dichlorobenzene	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	1,4-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery
0608036-01	VOCs	2-Butanone (MEK)	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-01	VOCs	2-Butanone (MEK)	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	2-Butanone (MEK)	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-01	VOCs	Acetone	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-01	VOCs	Acetone	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Bromobenzene	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Bromobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery
0608036-01	VOCs	Bromoform	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-01	VOCs	Bromoform	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Dichlorodifluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.

0608036-01	VOCs	Dichlorodifluoromethane	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Dichlorodifluoromethane	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-01	VOCs	Trichlorofluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-01	VOCs	Trichlorofluoromethane	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Vinyl chloride	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-01	VOCs	Vinyl chloride	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	1,1-Dichloroethene	C1	The reported concentration for this analyte is below the quantitation limit.
0608036-02	VOCs	2-Butanone (MEK)	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-02	VOCs	2-Butanone (MEK)	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	2-Butanone (MEK)	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-02	VOCs	Acetone	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-02	VOCs	Acetone	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	Bromoform	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-02	VOCs	Bromoform	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	Dichlorodifluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-02	VOCs	Dichlorodifluoromethane	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	Dichlorodifluoromethane	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-02	VOCs	Trichlorofluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-02	VOCs	Trichlorofluoromethane	J	The reported result for this analyte should be considered an estimated value.

0608036-02	VOCs	Vinyl chloride	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-02	VOCs	Vinyl chloride	J	The reported result for this analyte should be considered an estimated value.
0608036-03	VOCs	2-Butanone (MEK)	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-03	VOCs	2-Butanone (MEK)	J	The reported result for this analyte should be considered an estimated value.
0608036-03	VOCs	2-Butanone (MEK)	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-03	VOCs	Acetone	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-03	VOCs	Acetone	J	The reported result for this analyte should be considered an estimated value.
0608036-03	VOCs	Bromoform	C3	The initial calibration for this analyte did not meet calibration criteria.
0608036-03	VOCs	Bromoform	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	1,1,1,2-Tetrachloroethane	U	This analyte was not detected.
0608036-01	VOCs	1,1,1-Trichloroethane	U	This analyte was not detected.
0608036-03	VOCs	Dichlorodifluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-03	VOCs	Dichlorodifluoromethane	J	The reported result for this analyte should be considered an estimated value.
0608036-03	VOCs	Dichlorodifluoromethane	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-03	VOCs	Trichlorofluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-03	VOCs	Trichlorofluoromethane	J	The reported result for this analyte should be considered an estimated value.
0608036-03	VOCs	Vinyl chloride	C4	The calibration verification check did not meet % difference criteria for this analyte.
0608036-03	VOCs	Vinyl chloride	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	2-Butanone (MEK)	C3	The initial calibration for this analyte did not meet calibration criteria.

B6H0091-BLK1	VOCs	2-Butanone (MEK)	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	2-Butanone (MEK)	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
0608036-01	VOCs	1,1,2,2-Tetrachloroethane	U	This analyte was not detected.
0608036-01	VOCs	1,1,2-Trichloroethane	U	This analyte was not detected.
0608036-01	VOCs	1,1-Dichloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Acetone	C3	The initial calibration for this analyte did not meet calibration criteria.
B6H0091-BLK1	VOCs	Acetone	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	Bromoform	C3	The initial calibration for this analyte did not meet calibration criteria.
B6H0091-BLK1	VOCs	Bromoform	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	Dichlorodifluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
B6H0091-BLK1	VOCs	Dichlorodifluoromethane	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	Dichlorodifluoromethane	Q3	The quantitation limit standard did not meet recovery criteria for this analyte.
B6H0091-BLK1	VOCs	Trichlorofluoromethane	C4	The calibration verification check did not meet % difference criteria for this analyte.
B6H0091-BLK1	VOCs	Trichlorofluoromethane	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	Vinyl chloride	C4	The calibration verification check did not meet % difference criteria for this analyte.
B6H0091-BLK1	VOCs	Vinyl chloride	J	The reported result for this analyte should be considered an estimated value.
B6H0091-BLK1	VOCs	Vinyl chloride	J	This analyte was not detected.
0608036-01	VOCs	1,1-Dichloropropene	U	

This analyte was not detected.

0608036-01	VOCs	1,2,3-Trichlorobenzene	U	
B6H0091-MS1	VOCs	1,1,2,2-Tetrachloroethane	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MS1	VOCs	1,3-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MS1	VOCs	1,4-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MS1	VOCs	Bromobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MSD1	VOCs	1,3-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MSD1	VOCs	1,4-Dichlorobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
B6H0091-MSD1	VOCs	Bromobenzene	Q4	The matrix spike and/or matrix spike duplicate associated with this sample did not meet recovery criteria for this analyte (see MS/MSD results for this batch in QC summary)
0608036-01	VOCs	1,2,3-Trichloropropane	U	This analyte was not detected.
0608036-01	VOCs	1,2,4-Trichlorobenzene	U	This analyte was not detected.
0608036-01	VOCs	1,2,4-Trimethylbenzene	U	This analyte was not detected.

				This analyte was not detected.
0608036-01	VOCs	1,2-Dibromo-3-chloropropane	U	This analyte was not detected.
0608036-01	VOCs	1,2-Dibromoethane (EDB)	U	This analyte was not detected.
0608036-01	VOCs	1,2-Dichlorobenzene	U	This analyte was not detected.
0608036-01	VOCs	1,2-Dichloroethane	U	This analyte was not detected.
0608036-01	VOCs	1,2-Dichloropropane	U	This analyte was not detected.
0608036-01	VOCs	1,3,5-Trimethylbenzene	U	This analyte was not detected.
0608036-01	VOCs	1,3-Dichlorobenzene	U	This analyte was not detected.
0608036-01	VOCs	1,3-Dichloropropane	U	This analyte was not detected.
0608036-01	VOCs	1,4-Dichlorobenzene	U	This analyte was not detected.
0608036-01	VOCs	2,2-Dichloropropane	U	This analyte was not detected.
0608036-01	VOCs	2-Butanone (MEK)	U	This analyte was not detected.
0608036-01	VOCs	2-Chlorotoluene	U	This analyte was not detected.
0608036-01	VOCs	4-Chlorotoluene	U	This analyte was not detected.
0608036-01	VOCs	Acetone	U	This analyte was not detected.
0608036-01	VOCs	Benzene	U	This analyte was not detected.
0608036-01	VOCs	Bromobenzene	U	This analyte was not detected.
0608036-01	VOCs	Bromochloromethane	U	This analyte was not detected.
0608036-01	VOCs	Bromodichloromethane	U	This analyte was not detected.

				This analyte was not detected.
0608036-01	VOCs	Bromoform	U	This analyte was not detected.
0608036-01	VOCs	Bromomethane	U	This analyte was not detected.
0608036-01	VOCs	Butylbenzene	U	This analyte was not detected.
0608036-01	VOCs	Carbon tetrachloride	U	This analyte was not detected.
0608036-01	VOCs	Chlorobenzene	U	This analyte was not detected.
0608036-01	VOCs	Chlorodibromomethane	U	This analyte was not detected.
0608036-01	VOCs	Chloroethane	U	This analyte was not detected.
0608036-01	VOCs	Chloroform	U	This analyte was not detected.
0608036-01	VOCs	Chloromethane	U	This analyte was not detected.
0608036-01	VOCs	cis-1,2-Dichloroethene	U	This analyte was not detected.
0608036-01	VOCs	cis-1,3-Dichloropropene	U	This analyte was not detected.
0608036-01	VOCs	Dibromomethane	U	This analyte was not detected.
0608036-01	VOCs	Dichlorodifluoromethane	U	This analyte was not detected.
0608036-01	VOCs	Dichloromethane	U	This analyte was not detected.
0608036-01	VOCs	Ethylbenzene	U	This analyte was not detected.
0608036-02	VOCs	1,1,2-Trichloro-1,2,2-trifluoroethane(U	This analyte was not detected.
0608036-02	VOCs	1,1,2-Trichloroethane	U	This analyte was not detected.
0608036-02	VOCs	1,1-Dichloroethane	U	This analyte was not detected.

0608036-02	VOCs	1,1-Dichloropropene	U	This analyte was not detected.
0608036-02	VOCs	1,2,3-Trichlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	1,2,3-Trichloropropane	U	This analyte was not detected.
0608036-02	VOCs	1,2,4-Trichlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	1,2,4-Trimethylbenzene	U	This analyte was not detected.
0608036-02	VOCs	1,2-Dibromo-3-chloropropane	U	This analyte was not detected.
0608036-02	VOCs	1,2-Dibromoethane (EDB)	U	This analyte was not detected.
0608036-02	VOCs	1,2-Dichlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	1,2-Dichloroethane	U	This analyte was not detected.
0608036-02	VOCs	1,2-Dichloropropane	U	This analyte was not detected.
0608036-02	VOCs	1,3,5-Trimethylbenzene	U	This analyte was not detected.
0608036-02	VOCs	1,3-Dichlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	1,3-Dichloropropane	U	This analyte was not detected.
0608036-02	VOCs	1,4-Dichlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	2,2-Dichloropropane	U	This analyte was not detected.
0608036-02	VOCs	2-Butanone (MEK)	U	This analyte was not detected.
0608036-02	VOCs	2-Chlorotoluene	U	This analyte was not detected.
0608036-02	VOCs	4-Chlorotoluene	U	This analyte was not detected.

				This analyte was not detected.
0608036-02	VOCs	Acetone	U	This analyte was not detected.
0608036-02	VOCs	Benzene	U	This analyte was not detected.
0608036-02	VOCs	Bromobenzene	U	This analyte was not detected.
0608036-02	VOCs	Bromochloromethane	U	This analyte was not detected.
0608036-02	VOCs	Bromodichloromethane	U	This analyte was not detected.
0608036-02	VOCs	Bromoform	U	This analyte was not detected.
0608036-02	VOCs	Bromomethane	U	This analyte was not detected.
0608036-02	VOCs	Butylbenzene	U	This analyte was not detected.
0608036-02	VOCs	Carbon tetrachloride	U	This analyte was not detected.
0608036-02	VOCs	Chlorobenzene	U	This analyte was not detected.
0608036-02	VOCs	Chlorodibromomethane	U	This analyte was not detected.
0608036-02	VOCs	Chloroethane	U	This analyte was not detected.
0608036-02	VOCs	Chloroform	U	This analyte was not detected.
0608036-02	VOCs	Chloromethane	U	This analyte was not detected.
0608036-02	VOCs	cis-1,2-Dichloroethene	U	This analyte was not detected.
0608036-02	VOCs	cis-1,3-Dichloropropene	U	This analyte was not detected.
0608036-02	VOCs	Dibromomethane	U	This analyte was not detected.
0608036-02	VOCs	Dichlorodifluoromethane	U	This analyte was not detected.

				This analyte was not detected.
0608036-02	VOCs	Dichloromethane	U	This analyte was not detected.
0608036-02	VOCs	Ethylbenzene	U	This analyte was not detected.
0608036-02	VOCs	Hexachlorobutadiene	U	This analyte was not detected.
0608036-02	VOCs	Isopropylbenzene	U	This analyte was not detected.
0608036-02	VOCs	m&p-Xylene	U	This analyte was not detected.
0608036-02	VOCs	Naphthalene	U	This analyte was not detected.
0608036-02	VOCs	o-Xylene	U	This analyte was not detected.
0608036-02	VOCs	p-Isopropyltoluene	U	This analyte was not detected.
0608036-02	VOCs	Propylbenzene	U	This analyte was not detected.
0608036-02	VOCs	sec-Butylbenzene	U	This analyte was not detected.
0608036-02	VOCs	Styrene	U	This analyte was not detected.
0608036-02	VOCs	tert-Butyl methyl ether (MTBE)	U	This analyte was not detected.
0608036-02	VOCs	tert-Butylbenzene	U	This analyte was not detected.
0608036-02	VOCs	Toluene	U	This analyte was not detected.
0608036-02	VOCs	trans-1,2-Dichloroethene	U	This analyte was not detected.
0608036-02	VOCs	trans-1,3-Dichloropropene	U	This analyte was not detected.
0608036-02	VOCs	Trichloroethene	U	This analyte was not detected.
0608036-02	VOCs	Trichlorofluoromethane	U	This analyte was not detected.

0608036-02	VOCs	Vinyl chloride	U	This analyte was not detected.
0608036-03	VOCs	1,1,1,2-Tetrachloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,1,1-Trichloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,1,2,2-Tetrachloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,1,2-Trichloro-1,2,2-trifluoroethane(U	This analyte was not detected.
0608036-03	VOCs	1,1,2-Trichloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,1-Dichloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,1-Dichloroethene	U	This analyte was not detected.
0608036-03	VOCs	1,1-Dichloropropene	U	This analyte was not detected.
0608036-03	VOCs	1,2,3-Trichlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	1,2,3-Trichloropropane	U	This analyte was not detected.
0608036-03	VOCs	1,2,4-Trichlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	1,2,4-Trimethylbenzene	U	This analyte was not detected.
0608036-03	VOCs	Bromodichloromethane	U	This analyte was not detected.
0608036-03	VOCs	Bromoform	U	This analyte was not detected.
0608036-03	VOCs	Bromomethane	U	This analyte was not detected.
0608036-03	VOCs	Butylbenzene	U	This analyte was not detected.
0608036-03	VOCs	Carbon tetrachloride	U	This analyte was not detected.

				This analyte was not detected.
0608036-03	VOCs	Chlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	Chlorodibromomethane	U	This analyte was not detected.
0608036-03	VOCs	Chloroethane	U	This analyte was not detected.
0608036-03	VOCs	Chloroform	U	This analyte was not detected.
0608036-03	VOCs	Chloromethane	U	This analyte was not detected.
0608036-03	VOCs	cis-1,2-Dichloroethene	U	This analyte was not detected.
0608036-03	VOCs	cis-1,3-Dichloropropene	U	This analyte was not detected.
0608036-03	VOCs	Dibromomethane	U	This analyte was not detected.
0608036-03	VOCs	Dichlorodifluoromethane	U	This analyte was not detected.
0608036-03	VOCs	Dichloromethane	U	This analyte was not detected.
0608036-03	VOCs	Ethylbenzene	U	This analyte was not detected.
0608036-03	VOCs	Hexachlorobutadiene	U	This analyte was not detected.
0608036-03	VOCs	Isopropylbenzene	U	This analyte was not detected.
0608036-03	VOCs	m&p-Xylene	U	This analyte was not detected.
0608036-03	VOCs	Naphthalene	U	This analyte was not detected.
0608036-03	VOCs	o-Xylene	U	This analyte was not detected.
0608036-03	VOCs	p-Isopropyltoluene	U	This analyte was not detected.
0608036-03	VOCs	Propylbenzene	U	This analyte was not detected.

0608036-03	VOCs	sec-Butylbenzene	U	This analyte was not detected.
0608036-03	VOCs	Styrene	U	This analyte was not detected.
0608036-03	VOCs	tert-Butyl methyl ether (MTBE)	U	This analyte was not detected.
0608036-03	VOCs	tert-Butylbenzene	U	This analyte was not detected.
0608036-03	VOCs	Toluene	U	This analyte was not detected.
0608036-03	VOCs	trans-1,2-Dichloroethene	U	This analyte was not detected.
0608036-03	VOCs	trans-1,3-Dichloropropene	U	This analyte was not detected.
0608036-03	VOCs	Trichloroethene	U	This analyte was not detected.
0608036-03	VOCs	Trichlorofluoromethane	U	This analyte was not detected.
0608036-03	VOCs	Vinyl chloride	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1,1,2-Tetrachloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1,1-Trichloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1,2,2-Tetrachloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1,2-Trichloro-1,2,2-trifluoroethane(U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1,2-Trichloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1-Dichloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1-Dichloroethene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,1-Dichloropropene	U	This analyte was not detected.

B6H0091-BLK1	VOCs	1,2,3-Trichlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2,3-Trichloropropane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2,4-Trichlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2,4-Trimethylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2-Dibromo-3-chloropropane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2-Dibromoethane (EDB)	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2-Dichlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2-Dichloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,2-Dichloropropane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,3,5-Trimethylbenzene	U	This analyte was not detected.
0608036-01	VOCs	Hexachlorobutadiene	U	This analyte was not detected.
0608036-01	VOCs	Isopropylbenzene	U	This analyte was not detected.
0608036-01	VOCs	m&p-Xylene	U	This analyte was not detected.
0608036-01	VOCs	Naphthalene	U	This analyte was not detected.
0608036-01	VOCs	o-Xylene	U	This analyte was not detected.
0608036-01	VOCs	p-Isopropyltoluene	U	This analyte was not detected.
0608036-01	VOCs	Propylbenzene	U	This analyte was not detected.
0608036-01	VOCs	sec-Butylbenzene	U	This analyte was not detected.

0608036-01	VOCs	Styrene	U	This analyte was not detected.
0608036-01	VOCs	tert-Butyl methyl ether (MTBE)	U	This analyte was not detected.
0608036-01	VOCs	tert-Butylbenzene	U	This analyte was not detected.
0608036-01	VOCs	Toluene	U	This analyte was not detected.
0608036-01	VOCs	trans-1,2-Dichloroethene	U	This analyte was not detected.
0608036-01	VOCs	trans-1,3-Dichloropropene	U	This analyte was not detected.
0608036-01	VOCs	Vinyl chloride	U	This analyte was not detected.
0608036-02	VOCs	1,1,1,2-Tetrachloroethane	U	This analyte was not detected.
0608036-02	VOCs	1,1,1-Trichloroethane	U	This analyte was not detected.
0608036-02	VOCs	1,1,2,2-Tetrachloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,2-Dibromo-3-chloropropane	U	This analyte was not detected.
0608036-03	VOCs	1,2-Dibromoethane (EDB)	U	This analyte was not detected.
0608036-03	VOCs	1,2-Dichlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	1,2-Dichloroethane	U	This analyte was not detected.
0608036-03	VOCs	1,2-Dichloropropane	U	This analyte was not detected.
0608036-03	VOCs	1,3,5-Trimethylbenzene	U	This analyte was not detected.
0608036-03	VOCs	1,3-Dichlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	1,3-Dichloropropane	U	This analyte was not detected.

0608036-03	VOCs	1,4-Dichlorobenzene	U	This analyte was not detected.
0608036-03	VOCs	2,2-Dichloropropane	U	This analyte was not detected.
0608036-03	VOCs	2-Butanone (MEK)	U	This analyte was not detected.
0608036-03	VOCs	2-Chlorotoluene	U	This analyte was not detected.
0608036-03	VOCs	4-Chlorotoluene	U	This analyte was not detected.
0608036-03	VOCs	Acetone	U	This analyte was not detected.
0608036-03	VOCs	Benzene	U	This analyte was not detected.
0608036-03	VOCs	Bromobenzene	U	This analyte was not detected.
0608036-03	VOCs	Bromochloromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,3-Dichlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,3-Dichloropropane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	1,4-Dichlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	2,2-Dichloropropane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	2-Butanone (MEK)	U	This analyte was not detected.
B6H0091-BLK1	VOCs	2-Chlorotoluene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	4-Chlorotoluene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Acetone	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Benzene	U	This analyte was not detected.

				This analyte was not detected.
B6H0091-BLK1	VOCs	Bromobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Bromochloromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Bromodichloromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Bromoform	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Bromomethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Butylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Carbon tetrachloride	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Chlorobenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Chlorodibromomethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Chloroethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Chloroform	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Chloromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	cis-1,2-Dichloroethene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	cis-1,3-Dichloropropene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Dibromomethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Dichlorodifluoromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Dichloromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Ethylbenzene	U	This analyte was not detected.

				This analyte was not detected.
B6H0091-BLK1	VOCs	Hexachlorobutadiene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Isopropylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	m&p-Xylene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Naphthalene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	o-Xylene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	p-Isopropyltoluene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Propylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	sec-Butylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Styrene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	tert-Butyl methyl ether (MTBE)	U	This analyte was not detected.
B6H0091-BLK1	VOCs	tert-Butylbenzene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Tetrachloroethene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Toluene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	trans-1,2-Dichloroethene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	trans-1,3-Dichloropropene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Trichloroethene	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Trichlorofluoromethane	U	This analyte was not detected.
B6H0091-BLK1	VOCs	Vinyl chloride	U	This analyte was not detected.

0608036-01	VOCs	1,1-Dichloroethene	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Ethane, -dichloro-trifluo	J	The reported result for this analyte should be considered an estimated value.
0608036-01	VOCs	Methane, dichlorofluoro-	J	The reported result for this analyte should be considered an estimated value.
0608036-02	VOCs	1,1-Dichloroethene	J	The reported result for this analyte should be considered an estimated value.